

Qualification Report

Combined FutMon/ICP-Forest Water Ring Test 2010

Laboratory

D.Deneva and S.Bratinova
Executive Environment Agency
Water Quality Laboratory
1618 - Sofia
Bulgaria

deneva@nfp-bg.eionet.eu.int

Associated Beneficiary

Genoveva Popova
Executive Agency, Monitoring of Land, Biodiversity and
Ministry of Environment and Water, Environmental
136, Tzar Boris III blvd.
1618 SOFIA

forest@nfp-bg.eionet.eu.int

Labcode: D05

Parameter	analysed	passed	not passed	Percentage of correct results	Remarks	After requalification
NH ₄	yes	X		100%		
Ca	yes	X		100%		
Cl	yes	X		100%		
Conductivity	yes	X		100%		
Mg	yes	X		100%		
N-NO ₃	yes	X		100%		
pH	yes	X		100%		
K	yes	X		100%		
Na	yes	X		100%		
S-SO ₄	yes	X		100%		
Total N	yes	X		80%		
DOC	yes	X		80%		
Alkalinity	yes	X		50%		

passed: 50% of the results per element or more are within the tolerable limits

not passed: less than 50% of the results per element are within the tolerable limits

The results meet the WRT overall data quality objectives.

24/05/2010



Aldo Marchetto
Leader of FutMon Action C1-Water



LIFE07 ENV/D/000218